

# BIOCOMPATIBILITY CERTIFICATE

**Material tested:** SILPURAN® 6000/50, WE 08930

**Manufacturer:** WACKER Chemie AG, Germany  
WACKER Chemical Corporation, USA  
WACKER Asahikasei Silicone Co. Ltd., Japan

**Studies performed:** The following studies were performed in order to determine the biocompatibility of the test item according to ISO 10993-1:

**CYTOTOXICITY**  
**SENSITIZATION (LLNA)**  
**TEST FOR PYROGENS**

**Results:** The product did not show any adverse effects in the studies performed.

**BSL BIOSERVICE Scientific Laboratories GmbH**

Behringstraße 6/8

D-82152 Planegg



Dr. Daniela Stelter  
Biological Safety Testing  
Date: 30 October 2008

